


Application Guide

STEP

1


Open up the Elbowlift so that the hook-and-loop strap is pulled away from the suspension pad.



STEP

2


Place tip of patient's elbow over the centering hole with the foam insert pads aligned under the arm and forearm.



STEP

3

Close the strap across the elbow using the hook and loop closure. Fit the Elbowlift so it is snug but not tight.



Care of Elbowlift:

When Elbowlift becomes soiled, hand washing is recommended.

Machine washing Elbowlift in a standard washer is an acceptable alternative to hand washing. The hook-and-loop strap should be closed and placed into a net laundry bag prior to machine washing.

IMPORTANT: Remove Elbowlift every eight hours and check the skin or wound status.

For the prevention or treatment of heel ulcers, use



and



Suspension Boots.